

Specifications

OPTICAL TEMPERATURE SENSORS SPOTS & FOILS, RESPIRATION VIALS and FLOW-THROUGH CELLS

1 SENSOR SPECIFICATIONS

Specifications are valid for temperature sensor spots (item no.: **TPSP5**), self-adhesive temperature sensor spots (item no.: **TPSP5-ADH**), temperature sensor foil (item no.: **TPFOIL**), for the temperature sensor of respiration vials (item no.: **TOVIAL20**, **PHTOVIAL20-PKx**, **PHTVIAL20-PKx**), and for the temperature sensor of flow-through cells (item no.: **TPFTC2**, **TOFTC2**, **TPFLOW**).

1.1 Specifications in gas & water/aqueous samples

Specifications	
Measuring Range	0 to 50°C (273-323 K; 32-122 °F)
Accuracy	± 0.5°C (1-point calibrated)
Resolution	0.04°C
Drift	<0.2°C / week at 25°C

* For 1-point calibrated sensor.

1.2 General Characteristics

Response time ‡ (t90) in water	<1 s
Minimum Lifetime	10,000,000 data points
Calibration Modes	1-point calibration
Application Areas	Laboratory, industry, research. NOT for medical or any safety-critical application. NOT for application in humans. NOT for application in food intended for human consumption.

‡ Typical response times for 90% signal change.

2 APPLICABILITY AND CROSS-SENSITIVITY

	Applicability	Cross-Sensitivity
Water/Aqueous solutions	X	
Gas Phase (typ. air components)	X	
Ethanol ¹	short-term only	
Methanol ¹	short-term only	
Isopropanol ¹	short-term only	
Other organic solvents ²		X
Bleach		X
pH 1-14 ³	X	

¹ Only for temperature compensation of oxygen measurements (in combination with contactless oxygen sensor).

² Includes liquid solvents and solvent vapors

³ pH 2-9 for **TPSP5-ADH**

3 CLEANING, STERILIZATION, STORAGE

Cleaning	3% H ₂ O ₂ , Soap solution, short-term Ethanol
Sterilization	<p>TPSP5: Sterilization with short-term 70% Ethanol and 70% Isopropanol; autoclavable few cycles at 121°C for 15 min with special precautions (details on request).</p> <p>TOVIAL20: Sterilization with short-term 70% Ethanol and 70% Isopropanol; only without lid & septum autoclavable few cycles at 121°C for 15 min with special precautions (details on request).</p> <p>PHTOVIAL20-PKx, PHTVIAL20-PKx, TPFTC2, TOFTC2: Sensor vials with integrated pH sensors and all flow-through cells are NOT autoclavable.</p> <p>TPFLOW: delivered pre-sterilized with 25kGy beta-radiation, Sterilization with short-term 70% Ethanol and 70% Isopropanol; autoclavable few cycles at 121°C for 15 min without the optical fiber.</p>
Storage	>3 years in darkness at room temperature

Contact

PyroScience GmbH
 Hubertusstraße 35
 52064 Aachen
 Deutschland

Tel.: +49 (0)241 5183 2210
 Fax: +49 (0)241 5183 2299
 info@pyroscience.com
 www.pyroscience.com